

Certificate

VITO
oil filter system



Module type:

VITO[®] Oiltester

(Testo 270)

Serial no.:

20313871

Function check passed

Display-Test

LED-Functions

Push button test

Optic check passed

Damages

Sensor



Measured Values

Reference	Permissible Tolerance	Actual Value	Result
4,6% TPM	± 2,0% TPM	4,1% TPM	ok
50,3° C	±1,5° C	50,3° C	ok
26,0% TPM	± 2,0% TPM	25,5% TPM	ok
180,1° C	±1,5° C	180,4° C	ok

Date:

22.09.2014

Valid until:

22.09.2015

Inspector:

Sascha Baisch

